

EUROLINE Atopy “South East Asia 3” (IgE) Test instruction











ORDER NO.	ANTIBODIES AGAINST	IG CLASS	SUBSTRATE	FORMAT
DP 3715-1601-3 E	Inhalation and food allergens	IgE	Test strips coated with allergens	16 x 01 (16)

Indication: The EUROLINE test kit supports the diagnosis of sensitisations leading to allergy-associated symptoms, e.g. conjunctivitis, rhinitis or gastrointestinal problems by detection of specific IgE.

Application: The EUROLINE test kit provides semiquantitative in vitro determination of allergen-specific IgE (sIgE) in serum or plasma, contributing to the diagnosis of allergies. The test is a multiparameter assay containing optimised combinations of relevant allergens, enabling the analysis of sIgE against these different allergens in one test.

Principle of the test: The test kit contains test strips coated with 21 different allergens. The test strips are first moistened and then incubated with patient samples in the first reaction step. If samples contain specific antibodies of class IgE, they will bind to the allergens coated on the strip. To detect the bound antibodies, a second incubation is carried out using an enzyme-labelled anti-human IgE (enzyme conjugate) catalysing a colour reaction.

Contents of a test kit:

Description	Format	Symbol
1. Test strips coated with the allergens ts20, w1, w6, ds1, h1, e1, e2, i6, ms1, u80, f1, f2, f13, f14, f27, f88, fs33, fs34, f24, f23, CCD	16 strips	
2. Enzyme conjugate Alkaline phosphatase-labelled anti-human IgE (mouse), ready for use	1 x 20 ml	
3. Universal buffer 10x concentrated	1 x 100 ml	
4. Substrate solution Nitroblue tetrazolium chloride/5-Bromo-4-chloro-3-indolylphosphate (NBT/BCIP), ready for use	1 x 30 ml	
5. Incubation tray , volume-reduced (400 µl)	2 x 10 channels	
6. Plastic foil	1 sheet	
7. Instruction booklet	1 booklet	
 Lot description		 Storage temperature
 In vitro diagnostic medical device		 Unopened usable until

Performance of the test requires incubation trays or other components, which are not provided in the test kits. They are available from EUROIMMUN under the following order numbers:

ZD 9897-0130 Incubation tray (volume-reduced 400 µl) with 30 channels (black, compatible with EUROBlotMaster and EUROBlotCamera System)

ZD 9897-0144 Incubation tray (volume-reduced 400 µl) with 44 channels (black, compatible with EUROBlotOne, EUROBlotMaster and EUROBlotCamera System)

ZD 9895-0130 Incubation tray (volume 1 ml) with 30 channels (black, compatible with EUROBlotMaster and EUROBlotCamera System)

ZD 9898-0144 Incubation tray (volume 1 ml) with 44 channels (black, compatible with EUROBlotOne, EUROBlotMaster and EUROBlotCamera System)

Modifications to the former version are marked in grey.



For the creation of work protocols and the evaluation of incubated test strips using **EUROLineScan** you require green paper and adhesive foil:

ZD 9880-0101 Green paper (1 sheet)

ZD 9885-0116 Adhesive foil for approx. 16 test strips

ZD 9885-0130 Adhesive foil for approx. 30 test strips

For covering the incubation trays the adhesive foil can be used as well.

Preparation and stability of the reagents

Note: All reagents must be brought to room temperature (+18°C to +25°C) approx. 30 minutes before use. Unopened, reagents are stable until the indicated expiry date when stored at +2°C to +8°C. After initial opening, reagents are stable for 12 months or until the expiry date, unless stated otherwise below. Opened reagents must also be stored at +2°C to +8°C and protected from contamination.

- **Coated test strips:** Ready for use. Open the package with the test strips only when the strips have reached room temperature (+18°C to +25°C) to prevent condensation on the strips. After removal of the strips the package should be sealed tightly and stored at +2°C to +8°C.
- **Enzyme conjugate:** Ready for use. Mix thoroughly before using.
- **Universal buffer:** The universal buffer is supplied as a 10x concentrate. For the preparation of the working strength universal buffer shake the bottle. The amount required should be removed from the bottle using a clean pipette and diluted 1:10 with deionised or distilled water. Due to the special membrane used for the present EUROLINE the working strength universal buffer is used for the dilution of patient samples and the washing of the test strips. For the incubation of 1 test strip 2.0 ml buffer concentrate should be diluted with 18.0 ml water. The working strength universal buffer should be used on the same working day.
- **Substrate solution:** Ready for use. Close bottle immediately after use, as the contents are sensitive to light ☀.

Storage and stability: The test kit must be stored at a temperature between +2°C and +8°C. Do not freeze. Unopened, all test kit components are stable until the indicated expiry date.

Waste disposal: Patient samples and incubated test strips should be handled as infectious waste. Other reagents do not need to be collected separately, unless stated otherwise in official regulations.

Warning: Some of the reagents contain sodium azide in a non-declarable concentration. Avoid skin contact.

Preparation and stability of the patient samples

Sample material: Human serum or EDTA, heparin or citrate plasma.

Stability: Patient samples to be investigated can generally be stored at +2°C to +8°C for up to 14 days. Diluted samples should be incubated within one working day.

Sample dilution:

Version a: The patient samples to be investigated are used undiluted.

Version b: Dilute 175 µl patient sample with 250 µl working strength universal buffer and mix thoroughly by vortexing. The final volume should be 425 µl.

Version c: Dilute 100 µl patient sample with 1.0 ml working strength universal buffer and mix thoroughly by vortexing. The final volume should be 1.1 ml.

Sample pipettes are not suitable for mixing.




Incubation

- Pretreat:** Place the required amount of test strips in the incubation tray. Fill each of the channels with 1.0 ml working strength universal buffer and incubate the test strips for **5 minutes**. Afterwards aspirate off all the liquid.
- Sample incubation:** (1st step)
- Manual:**
- Version a** (time-optimised): Fill each channel of the volume-reduced incubation tray with **400 µl** of **undiluted sample** and incubate for **60 minutes** at room temperature (+18°C to +25°C) on a rocking shaker.
- Version b** (volume-/time optimised): Fill each channel of the volume-reduced incubation tray with **425 µl** of **diluted sample** (175 µl sample + 250 µl working strength universal buffer) and incubate for **2 h** at room temperature (+18°C to +25°C) on a rocking shaker.
- Version c** (volume-optimised): Fill each channel with **1.1 ml** of **1:11 diluted sample** and incubate **overnight (12 to 24 h)** on a rocking shaker at room temperature (+18°C to +25°C). The use of incubation trays with capacities of 1 ml and 400 µl is possible.
(Cover the incubation tray to prevent evaporation.)
- Automatic:**
- Version b** (volume-/time-optimised): The incubation volume must be increased to 510 µl (210 µl sample + 300 µl working strength universal buffer).
- Version c** (volume-optimised): In version c with 1:11 diluted sample overnight, the incubation volume must be increased to 1.65 ml (150 µl sample + 1.5 ml working strength universal buffer).
- Caution:** Version c is not suitable for volume-reduced trays, 1 ml trays should be used.
- Washing:**
- Manual:**
Aspirate off the liquid from each channel and wash for **3 x 5 minutes** with 1.0 ml working strength universal buffer on a rocking shaker.
- Automatic:**
In version c with 1:11 diluted sample overnight, a volume of 1800 µl must be used for the first wash step.
- Conjugate incubation:** (2nd step) Pipette 1.0 ml enzyme conjugate (alkaline phosphatase-conjugated anti-human IgE) into each channel.
Incubate for **60 minutes** at room temperature (+18°C to +25°C) on a rocking shaker.
- Washing:** Aspirate off the liquid from each channel. Wash as described above.
- Substrate incubation:** (3rd step) Pipette 1.0 ml substrate solution into each channel.
Incubate for **10 minutes** at room temperature (+18°C to +25°C) on a rocking shaker.
- Stopping:** Aspirate off the liquid from each channel and wash each strip **3 x 1 minute** with deionised or distilled water.
- Evaluate:** Place the test strip on the evaluation protocol, air dry and evaluate.
- For automated incubation with the **EUROBlotMaster** select the program **Euro11 Allerg EL60** (version a), **Euro08 Allerg 2h** (version b) or **Euro12 Allerg 16h** (version c).
- For automated incubation with the **EUROBlotOne** select the program **EURO 11 Allergy EL60** (version a), **EURO 08 Allergy 2h** (version b) or **EURO 12 Allergy 16h** (version c).



EUROLINE Atopy "South East Asia 3" (IgE)

Incubation protocol





Interpretation of results

Handling: After stopping the reaction using deionised or distilled water, place the incubated test strips onto the adhesive foil of the green work protocol (created beforehand in the EUROLinScan program) using a pair of tweezers. The position of the test strips can be corrected while they are wet. As soon as all test strips have been placed onto the protocol, they should be pressed hard using filter paper and left to air-dry. The drying process should take place without any direct light, in an environment as dark as possible. After they have dried, the test strips will be stuck to the adhesive foil. Incubated strips that are still moist show a background colouring that disappears when they are completely dry. Therefore the evaluation of the strips is only to take place after the strips have completely dried.

For digital evaluation follow the instructions in the EUROLinScan user manual. The code for entering the **test** into EUROLinScan is **Atopy South East Asia 3**.

Caution: An indicator band is located on the lower end of the strip. The incubation was performed correctly if a colour reaction is visible on the control band. The test is valid if the indicator is detected at least with EAST class 3.

Some samples might display a dark background staining of the membrane and white bands at the position of the antigens. These lighter bands should be interpreted as negative.

An indicator band < EAST class 3 indicates an incorrect incubation. The incubation should be repeated with new reagents.

When using EUROLinScan the intensity of the bands is calculated in EAST classes of 0 to 6. EAST is the abbreviation for Enzyme-Allergo-Sorbent Test and is with respect to the concentration grades identical to the well-known RAST system (Radio-Allergo-Sorbent Test) used in allergy diagnostics.


The classes can be divided into the following concentrations:

Class	Concentration [kU/l]	Result
0	< 0.35	No specific antibodies detected.
1	$0.35 \leq \text{slgE} < 0.7$	Very low antibody titer, frequently no clinical symptoms where sensitisation is present.
2	$0.7 \leq \text{slgE} < 3.5$	Low antibody titer, existing sensitisation, frequently with clinical symptoms in the upper range of class.
3	$3.5 \leq \text{slgE} < 17.5$	Significant antibody titer, clinical symptoms usually present.
4	$17.5 \leq \text{slgE} < 50.0$	High antibody titer, almost always with clinical symptoms.
5	$50.0 \leq \text{slgE} < 100.0$	Very high antibody titer.
6	≥ 100.0	Very high antibody titer.

For diagnosis, the clinical picture of the patient always needs to be taken into account along with the serological findings.



The test strips include the following allergens:



Position	Allergen code	Allergen name
1	ts20	Tree mix 2 (t8, t12 and t14)
2	w1	Common ragweed
3	w6	Mugwort
4	ds1	House dust mite mix 1 (d1 and d2)
5	h1	House dust
6	e1	Cat
7	e2	Dog
8	i6	Cockroach, German
9	ms1	Mould mix 1 (m1, m2, m3 and m6)
10	u80	Hop
11	f1	Egg white
12	f2	Cow's milk
13	f13	Peanut
14	f14	Soybean
15	f27	Beef
16	f88	Lamb meat
17	fs33	Sea fish mix 1 (Codfish, Lobster and Scallop)
18	fs34	Fresh water fish mix (Salmon, Perch and Carp)
19	f24	Shrimp
20	f23	Crab
21	CCD	CCD marker
a	Ind	Indicator band

Test characteristics

Measurement range: The EUROLINE is a semiquantitative method. The measurement range is given in EAST system classes 0 to 6.

Cross reactions: Due to the similar structure of the allergens, e.g. similarities in chemical substances or botanical relations, cross reactions may occur. The specific IgE antibodies that have developed in a patient also attach to identical epitopes of homologous protein allergens.

Examples of cross reactivity between airborne allergens and food allergens:

Inhalation allergens	Associated food allergy
Grass	Tomato, potato, carrot, celery, garlic, onion, wheat, rice, green pea, peanut, apple, peach, orange, melon, kiwi
Birch	Hazelnut, walnut, apple, pear, carrot, celery, potato, orange, kiwi
Mugwort	Celery, carrot, spices, mustard, hazelnut
Ragweed	Melon, cucumber, banana
English plantain	Melon
Latex	Avocado, potato, banana, tomato, walnut, kiwi



Interference: Haemolytic, lipaemic and icteric sera up to a concentration of 5 mg/ml for haemoglobin, of 20 mg/ml for triglycerides and of 0.4 mg/ml for bilirubin showed no effect on the analytical results of the present EUROLINE.

Inter- and intra-assay variation: The inter-assay variation was determined by multiple analyses of characterised samples over several days. The intra-assay variation was determined by multiple analyses of characterised samples on one day. In every case, the intensity of the bands was within the specified range. This EUROLINE displays excellent inter- and intra-assay reproducibility.

Sensitivity and specificity: The sensitivity of the EUROLINE with respect to the ImmunoCAP system is for timothy grass (g6) 90%, for birch (t3) 90%, for Dermatophagoides pteronyssinus (d1) 83%, for Dermatophagoides farinae (d2) 84%, for cat (e1) 98% and for horse (e3) 82%.

The specificity of the EUROLINE with respect to the ImmunoCAP system is for timothy grass (g6) 100%, for birch (t3) 92%, for Dermatophagoides pteronyssinus (d1) 100%, for Dermatophagoides farinae (d2) 86%, for cat (e1) 91% and for horse (e3) 100%.

Timothy grass (g6)		EUROLINE	
	n = 94	pos.	neg.
	pos.	64	7
ImmunoCAP	neg.	0	23

Dermatoph. pt. (d1)		EUROLINE	
	n = 44	pos.	neg.
	pos.	25	5
ImmunoCAP	neg.	0	14

Cat (e1)		EUROLINE	
	n = 74	pos.	neg.
	pos.	50	1
ImmunoCAP	neg.	2	21

Birch (t3)		EUROLINE	
	n = 97	pos.	neg.
	pos.	55	6
ImmunoCAP	neg.	3	33

Dermatoph. far. (d2)		EUROLINE	
	n = 45	pos.	neg.
	pos.	26	5
ImmunoCAP	neg.	2	12

Horse (e3)		EUROLINE	
	n = 34	pos.	neg.
	pos.	9	2
ImmunoCAP	neg.	0	23

Limitations of in vitro allergy diagnostics

Accurate performance of the assays according to the test instruction will lead to reliable and reproducible results. In any case, the final diagnosis should not be solely based on one type of analysis. A well-founded anamnesis and further laboratory findings should always be taken into account. Skin test as well as provocation test (if possible) are mandatory to receive the entire information needed for an optimal decision regarding the specific immunotherapy that should be applied. The clinical picture is not always in line with in vitro test results.

Negative in vitro results may occur e.g. when:

- symptoms are not IgE-mediated,
- samples were taken before the organism was able to produce antibodies against the antigen,
- IgE concentrations reached a minimum a long time after sensitisation.

Positive results with specific IgE in vitro tests do not necessarily have to correlate with clinical manifestations.

Many IgE antibodies can cross-react with various allergens or redundant carbohydrate structures.

Especially food allergens frequently show a negative result in vitro although clinical symptoms may be present. This phenomenon can be explained through the effect of maturing, industrial processing, cooking, or frying of the allergen. Furthermore, the allergic reaction can be induced by a metabolite of the allergen resulting from the digestive process, which cannot exactly be recapitulated by in vitro diagnostics. Allergens bind to various degrees to the solid phase; this can influence the test results.

For the determination of specific IgE antibodies a variety of test systems is available. Due to the variability of the source material used for the production of allergen extracts and the manufacturing process itself the quality of the extracts used for allergy diagnostics varies significantly. Therefore the results of different test systems cannot easily be compared to each other due to the lack of international standards for both the allergens and the antibodies used in these assays. Thus, a slight deviation between different test systems cannot be ruled out.

In general, identical results for different patients do not necessarily mean identical clinical manifestations.



Clinical significance

The term allergy was defined in 1906 by the Austrian paediatrician Clemens von Pirquet to mean the body's increased ability to react to a foreign substance. Today allergy means an oversensitivity to foreign substances which are normally harmless. Alongside any genetic predisposition, numerous non-genetic factors also play a role, such as exposure to the allergen, nutritional condition, existing chronic diseases and acute viral infections. Atopy is a hereditary disposition to developing allergic reactions such as allergic asthma, rhinitis (hay fever) or dermatitis (including atopic eczema).

The most frequently occurring allergy is a type I hypersensitivity reaction, in which specific IgE antibodies are formed. The symptoms (rash, oedema or itching) generally occur shortly after contact with the allergen. These allergies are therefore also termed immediate type reactions. Allergens are acquired either through the air and mucous membranes of the body (inhalation allergies) or by ingestion (food allergies).

More than 15% of the population in industrial countries suffer from an immediate-type allergy. Typical allergic reactions are rhinitis, conjunctivitis and allergic asthma. A worldwide increase in allergic rhinitis has been observed, with a prevalence of 4% to more than 40% in various regions. Inhalation allergies can be triggered by seasonal allergens (pollen from trees, grasses and weeds) or all-year-round indoor allergens (house dust mites, domestic animals, mould spores). The allergic symptoms intensify with every further exposure to the allergen. If a systemic allergic reaction occurs, serious, even life-threatening reactions can result (anaphylactic shock).

A food allergy is an IgE mediated reaction which leads to symptoms within hours of having ingested the food. The most common foods causing allergic reactions are peanuts, soy, wheat, shellfish, fish, milk, eggs and tree nuts. Possible symptoms are burning or itching in the oral cavity, nausea, gastrointestinal spasms, diarrhoea and skin rashes. Severe reactions can also lead to asthma attacks, breathlessness, increased heart rate and to panic attacks and confusion. In rare cases anaphylaxis can occur (e.g. after consumption of peanuts, nuts or fish).

However, allergic reactions to foods of plant origin can also be caused by cross-reacting IgE antibodies. These reactions, termed cross-allergies, are based on the structural similarity of proteins which are present in both the food as well as in the corresponding inhalation allergens of plant origin. For example, patients with a birch pollen allergy can also develop allergic reactions to apple, celery, hazelnut, potato or kiwi.

Immunotherapy or hyposensitisation (desensitisation) established as a treatment for type I allergy does not provoke a change in IgE levels, although a significant reduction of the symptoms can be achieved. A definite response of a patient to immunotherapy normally manifests as an increase in the allergen-specific IgG antibody concentration in the course of treatment. However, this does not always correlate with a remission in symptoms.

Many allergens are glycoproteins and contain oligosaccharide side chains which are bound to the protein framework of the allergens. Some patients develop specific antibodies against these carbohydrate structures. The abbreviation CCD stands for cross-reactive carbohydrate determinant. CCDs are present in many plant and animal allergens. Due to their significant similarity in structure, CCDs are known to cause a strong cross reactivity. Although the importance of specific IgE antibodies against CCDs has not yet been fully understood, they are considered to be irrelevant for diagnosis in most cases and as such complicate the interpretation of positive in vitro diagnostic results. For this reason, the determination of specific IgE antibodies against CCDs may provide useful additional information, especially when positive IgE results disagree with the clinical picture, and can serve as an interpretation aid in the evaluation of overall test results.



Literature references

1. Arruda LK, Sole D, Baena-Cagnani CE, Nasпитz CK. **Risk factors for asthma and atopy.** Curr Opin Allergy Clin Immunol 5 (2005) 153-159.
2. Boullay ME, Boulet LP. **The relationships between atopy, rhinitis and asthma: pathophysiological considerations.** Curr Opin Allergy Clin Immunol 3 (2003) 51-55.
3. Caballero T, Martin-Esteban M. **Association between pollen hypersensitivity and edible vegetable allergy: A review.** Invest Allergol Clin Immunol 8 (1998) 6-16.
4. Eigenmann PA, Calza AM. **Diagnosis of IgE-mediated food allergy among Swiss children with atopic dermatitis.** Pediatr Allergy Immunol 11 (2000) 95-100.
5. EUROIMMUN AG. Meyer W, Scheper T, Lehmann H, Stöcker W. **Selbstklebende Blotmembranen.** Eingetragenes deutsches Gebrauchsmuster DE 202 15 268.5 (2003).
6. EUROIMMUN AG. Meyer W, Scheper T, Stöcker W. **Vorrichtung zur Antikörperdiagnose mit kombinierten Membranen.** Eingetragenes deutsches Gebrauchsmuster DE 202 15 270.7 (2003).
7. EUROIMMUN AG. Meyer W, Siegemund M, Stöcker W. **Vorrichtung zum immunenzymatischen Nachweis von Antikörpern in einer flüssigen Probe.** Eingetragenes deutsches Gebrauchsmuster DE 20 2006 006 622.5 (2006).
8. EUROIMMUN AG. Schlumberger W, Meyer W, Proost S, Dähnrich C, Müller-Kunert E, Sonnenberg K, Olbrich S, Stöcker W. **The new EUROBLot technology: Differentiation of Autoantibodies Against Cell Nuclei.** Eur J Clin Chem Biochem 33 (1995) 116.
9. EUROIMMUN AG. Stöcker W, Rateike M, Morrin M. **Verfahren zur Herstellung Festphasen-gebundener Bioreagenzien.** Europäische Patentanmeldung PCT/EP/2005/000974 (2005).
10. Frew AJ. **The immunology of respiratory allergies.** Toxicol Lett 86 (1996) 65-72.
11. Malandain H. **IgE-reactive carbohydrate epitopes-classification, cross-reactivity, and clinical impact.** Eur Ann Allergy Clin Immunol 37(4) (2005) 122-128.
12. Salkie ML. **Role of clinical laboratory in allergy testing.** Clin Biochem 27 (1994) 343-355.
13. Schoenwetter WF. **Allergic rhinitis: epidemiology and natural history.** Allergy Asthma Proc 21 (2000) 1-6.
14. Valenta R, Kraft D. **Type I allergic reactions to plant-derived food: A consequence of primary sensitization to pollen allergens.** J Allergy Clin Immunol 97 (1996) 893-895.





